# Maintenance Schedule for Aneroid Sphygmomanometers

**Each unit should be calibrated once each year, and every time the red part of the needle does not fall in the box located on the lower part of the dial. Please use the below table to record when each aneroid device is calibrated, and when each device is due for it's next calibration. Please clearly label each unit with a number for record-keeping purposes.

| Unit # and Model | Date Sent for Calibration | Next Scheduled Check | Name of Supervisor |
|---|---|---|---|
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |